

PAS calibration mode analysis V.2.0 (22nd Jan 2020)

The single real working calibration mode we have obtained during the SVT rehearsal with PAS EQM, DPU EM-2 at MSSL. This test has been done on 4th of June 2019. This text have been never executed with PFM onboard teh S/C with the updated DPU software.


The corresponding command is as follows:

|          |                             |     |     |    |          |               |        |
|----------|-----------------------------|-----|-----|----|----------|---------------|--------|
| ZIA58850 | SWA_TC_PAS_CALIBRATION_MODE | 203 | 128 | 10 | 0x0005C2 | Min HV CEM    | 1800 V |
|          |                             |     |     |    | 0x000066 | Step HV CEM   | 125 V  |
|          |                             |     |     |    | 0x00075B | Max HV CEM    | 2300 V |
|          |                             |     |     |    | 0x000FFB | Mask 0        | CEM #2 |
|          |                             |     |     |    | 0x000FF7 | Mask 1        | CEM #3 |
|          |                             |     |     |    | 0x000FEF | Mask 2        | CEM #4 |
|          |                             |     |     |    | 0x000DFD | Mask 3        | CEM #5 |
|          |                             |     |     |    | 0x000FBF | Mask 4        | CEM #6 |
|          |                             |     |     |    | 0x0006B7 | Normal CEM HV |        |
|          |                             |     |     |    | 0x000000 | Normal Mask   |        |

That my imagination about CEM mask was (0 is ON, 1 is OFF):

|   | 10 | 9 | 8 | 7 | 6 | 5 | 4 | 3 | 2                | 1               | 0              | CEM #          |
|---|----|---|---|---|---|---|---|---|------------------|-----------------|----------------|----------------|
| 1 | 1  | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1                | 1               | 0              | 11 FFFB CEM 2  |
| 1 | 1  | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1                | 1               | 0              | 111 FFF7 CEM 3 |
| 1 | 1  | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 1                | 0               | 111 FFEF CEM 4 |                |
| 1 | 1  | 1 | 1 | 1 | 1 | 1 | 1 | 1 | 0                | 1111 FFDF CEM 5 |                |                |
| 1 | 1  | 1 | 1 | 1 | 1 | 1 | 1 | 0 | 11111 FFBF CEM 6 |                 |                |                |

The general results are below, the cout was really at CEM #10. It means we cannot see the count in the "Calibration: packet.



The Calibration packed decoding is as follows (part)

```

1 | SCI.Parse_PAS_CAL | Final, start, step, acquisition = 1883 1474 102 25
1 | SCI.Parse_PAS_CAL | STEP = 0 ===== Start + step = 1474 E El CEM Count
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:13:56.500000000 Mask #1 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00000 : Acquisition 1/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:00.500000000 Mask #2 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00001 : Acquisition 2/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:04.500000000 Mask #3 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00002 : Acquisition 3/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:08.500000000 Mask #4 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00003 : Acquisition 4/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:12.500000000 Mask #5 37 4 10 112
1 | SCI.Parse_PAS_CAL | Record 00004 : Acquisition 5/25
1 | SCI.Parse_PAS_CAL | STEP = 1 ===== Start + step = 1576 E El CEM Count
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:20.500000000 Mask #1 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00005 : Acquisition 6/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:24.500000000 Mask #2 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00006 : Acquisition 7/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:28.500000000 Mask #3 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00007 : Acquisition 8/25
    
```

```

1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:32.500000000 Mask #4 0 0 0 0
1 | SCI.Parse_PAS_CAL | Record 00008 : Acquisition 9/25
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:36.500000000 Mask #5 37 4 10 112
1 | SCI.Parse_PAS_CAL | Record 00009 : Acquisition 10/25
1 | SCI.Parse_PAS_CAL | STEP = 2 ===== Start + step = 1678 E El CEM Count
1 | SCI.Parse_PAS_CAL | SCET = 2019-06-04T13:14:44.500000000 Mask #1 0 0 0 0
etc

```

The real TCL script used for the test on 4th Jun 2019 was as follows:

```

syslog "PAS enter Calibration mode"
tcsend ZIA58850 {PIA60780 0x0005C2} {PIA60781 0x000066} {PIA60782 0x00075B}
                {PIA60783 0x000FFB} {PIA60784 0x000FF7} {PIA60785 0x000FEF}
                {PIA60786 0x000DF0} {PIA60787 0x000BF0} {PIA60788 0x0006B7}
                {PIA60789 0x000000} checks {SPTV DPTV CEV} ack {ACCEPT COMPLETE}

```

The red parameters are incorrect, it seems it is an operator error.

The real TC contents of this day is as follows:

```

-----
APID (95,12) TC (203,128) SEQ (00046,3) Size = 22 : SWA_TC_PAS_CALIBRATION_MODE
0000 : 05 C2 00 66 07 5B 0F FB 0F F7 0F EF FD F0 FB F0
0016 : 06 B7 00 00 65 B7 CEM2 CEM3 CEM4 CEM9 CEM10
-----

```

Two last mask arguments correspond to CEM 9 and 10. So the TC conforms the observations.

Conclusions:

1. Data conform to the actual TC contents
2. We have to make a very short test with the part of SWA suite commissioning commands and test it again at MSSL.